



Bid Notice Abstract

Invitation to Bid (ITB)

Reference Number 10500503
Procuring Entity CARLOS HILADO MEMORIAL STATE UNIVERSITY
Title Procurement of Various Equipment of Capture Fisheries Laboratory for Instructional and Research of the College of Fisheries - Binalbagan Campus
Area of Delivery Negros Occidental

Solicitation Number:	CHMSU 24-011-0124-G	Status	Pending
Trade Agreement:	Implementing Rules and Regulations	Associated Components	7
Procurement Mode:	Public Bidding	Bid Supplements	0
Classification:	Goods	Document Request List	0
Category:	Laboratory Supplies and Equipment	Date Published	25/01/2024
Approved Budget for the Contract:	PHP 3,000,000.00	Last Updated / Time	24/01/2024 19:46 PM
Delivery Period:	60 Day/s	Closing Date / Time	15/02/2024 09:00 AM
Client Agency:			
Contact Person:	Ma. Lorena Fernandez Jugos Administrative Assistant II Mabini Street, Brgy Zone 1 Talisay City Negros Occidental Philippines 6115 63-34-7120005 Ext.142 63-939-9296624 bac.sec@chmsu.edu.ph		

Description

CARLOS HILADO MEMORIAL STATE UNIVERSITY
BIDS AND AWARDS COMMITTEE
TALISAY CITY, NEGROS OCCIDENTAL

INVITATION TO BID
FOR THE PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUCTIONAL AND RESEARCH OF THE COLLEGE OF FISHERIES - BINALBAGAN CAMPUS
CHMSU 24-011-0124-G

1. The Carlos Hilado Memorial State University, through the Corporate Budget for the Contract of 2024 approved by the governing Board (MDS-CO), intends to apply the sum of THREE MILLION PESOS & 00/100 (Php 3,000,000.00) ONLY, being the Approved Budget for the Contract (ABC) to payments under the contract for the PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUCTIONAL AND RESEARCH OF THE COLLEGE OF FISHERIES - BINALBAGAN CAMPUS. Bids received in excess of the ABC shall be automatically rejected at bid opening.

ITEM NO. DESCRIPTION ABC/ITEM
 1 ABSORBANCE/FLOURESCENCE SPECTROSCOPY 1,650,000.00
 2 ADVANCED DISSECTING MICROSCOPE 400,000.00
 3 AUTOCLAVE (STERILIZER) 200,000.00
 4 DIGITAL WATER FLOW METER 600,000.00
 5 GRAB TYPE BOTTOM SAMPLING DREDGE 150,000.00
 TOTAL ABC 3,000,000.00

2. The Carlos Hilado Memorial State University now invites bids for the above Procurement Project. Delivery of the Services is required by Sixty (60) calendar days after receipt of Notice to Proceed. Bidders should have completed, within two (2) years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II. Instructions to Bidders.

3. Bidding will be conducted through open competitive bidding procedures using the non-discretionary pass/fail criterion as specified in the 2016 Revised Implementing Rules and Regulations (IRR) of Republic Act 9184.

Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA 5183.

4. Prospective bidders may obtain further information from CARLOS HILADO MEMORIAL STATE UNIVERSITY and inspect the Bidding Documents at the address given below during 8:00 A.M. to 5:00 P.M.

The BAC Secretariat
Brgy. Zone 1, Mabini St.,
Talisay City, Negros Occidental

5. A complete set of Bidding Documents may be acquired by interested Bidders on January 25 – February 15, 2024 (8:30A.M.) from the given address and website(s) below and upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, in the maximum amount of Five Thousand Pesos (Php 5,000.00) Only. The Procuring Entity shall allow the bidder to present its proof of payment for the fees.

ITEM

NO.	DESCRIPTION	ABC	PER ITEM	COST OF BIDDING DOCUMENTS
1	ABSORBANCE/FLOURESCENCE SPECTROSCOPY	1,650,000.00	5,000.00	
2	ADVANCED DISSECTING MICROSCOPE	400,000.00	500.00	
3	AUTOCLAVE (STERILIZER)	200,000.00	500.00	
4	DIGITAL WATER FLOW METER	600,000.00	1,000.00	
5	GRAB TYPE BOTTOM SAMPLING DREDGE	150,000.00	500.00	

6. The CHMSU will hold a Pre-Bid conference on 9:30 A.M., February 1, 2024 at Bidding Room, 2/F Supply and Property Management Bldg., Brgy. Zone 1, Mabini St., Talisay City, Negros Occidental and or through video conferencing or webcasting via Zoom Meeting with ID No. 797 389 7583, Meeting Password 020124, which shall be open to prospective bidders.

7. Bids must be duly received by the BAC Secretariat through (i) manual submission at the Bidding Room, 2/F Supply and Property Management Bldg., Brgy. Zone 1, Mabini St., Talisay City, Negros Occidental, (ii) online or electronic submission at bac.sec@chmsu.edu.ph or (iii) both on or before 9:00 A.M., February 15, 2024. Late bids shall not be accepted.

Bids may be submitted through electronic mail to bac.sec@chmsu.edu.ph provided that the bidding documents are compressed into two (2) separate archived folders (zip or rar format) and which each folder shall be labelled as "First Envelope_Name of Company_Project Reference Number" and "Second Envelope_Name of Company_Project Reference Number" and each is uniquely password-protected;

Bidders must submit the printed copies (Original, Copy 1 and Copy 2) of their bidding documents within 3 calendar days from bid opening.

8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in ITB Clause 14.

9. Bid opening shall be on 9:30 A.M., February 15, 2024, at the Bidding Room, 2/F Supply and Property Management Bldg., Brgy. Zone 1, Mabini St., Talisay City, Negros Occidental through Zoom Meeting ID No. 797 389 7583, Meeting Password 021524. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.

10. The Carlos Hilado Memorial State University reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA 9184, without thereby incurring any liability to the affected bidder or bidders.

11. For further information, please refer to:

MS. LIGAYA E. FUENTES
Head-BAC Secretariat
Carlos Hilado Memorial State University
Brgy. Zone 1, Mabini St.,
Talisay City, Negros Occidental
Email Add.: bac.sec@chmsu.edu.ph
Telephone Nos.: (034) 454-0529; 454-0584 local 142
Mobile No.: 0939-9296624
Website: chmsu.edu.ph

12. You may visit the following websites:

For downloading of Bidding Documents: chmsu.edu.ph

For online bid submission: bac.sec@chmsu.edu.ph

ANDREW EUSEBIO S. TAN, Ph.D.
 BAC Chairperson

Line Items

Item No.	Product/Service Name	Description	Quantity	UOM	Budget (PHP)
1	ABSORBANCE/FLOURESCENE SPECTROSCOPY	.	1	Unit	1,650,000.00
2	ADVANCED DISSECTING MICROSCOPE	.	2	Unit	400,000.00
3	AUTOCLAVE (STERILIZER)	.	1	Unit	200,000.00
4	DIGITAL WATER FLOW METER	.	2	Unit	600,000.00
5	GRAB TYPE BOTTOM SAMPLING DREDGE	.	1	Unit	150,000.00

Pre-bid Conference

Date	Time	Venue
01/02/2024	9:30:00 AM	Bidding Room, 2/F Supply and Property Management Bldg., Brgy. Zone 1, Mabini St., Talisay City, Negros Occidental and or through video conferencing or webcasting via Zoom Meeting with ID No. 797 389 7583, Meeting Password 020124

Created by Rowena De la Vida Prado
Date Created 24/01/2024

The PhilGEPS team is not responsible for any typographical errors or misinformation presented in the system. PhilGEPS only displays information provided for by its clients, and any queries regarding the postings should be directed to the contact person/s of the concerned party.



Carlos Hilado Memorial State University

Alijis Campus • Binalbagan Campus • Fortune Towne Campus • Talisay (Main) Campus

A leading GREEN institution of higher learning in the global community

CARLOS HILADO MEMORIAL STATE UNIVERSITY
RECORDS MANAGEMENT OFFICE

RECEIVED

BY: [Signature]
DATE: JAN 25 2024 TIME: 10:50

Bids and Awards Committee

INVITATION TO BID FOR THE PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUCTIONAL AND RESEARCH OF THE COLLEGE OF FISHERIES – BINALBAGAN CAMPUS CHMSU 24-011-0124-G



- The *Carlos Hilado Memorial State University*, through the *Corporate Budget for the Contract of 2024 approved by the governing Board (MDS-CO)*, intends to apply the sum of **THREE MILLION PESOS & 007100 (Php 3,000,000.00) ONLY**, being the Approved Budget for the Contract (ABC) to payments under the contract for the **PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUCTIONAL AND RESEARCH OF THE COLLEGE OF FISHERIES - BINALBAGAN CAMPUS**. Bids received in excess of the ABC shall be automatically rejected at bid opening.

ITEM NO.	DESCRIPTION	ABC/ITEM
1	ABSORBANCE/FLOURESCENE SPECTROSCOPY	1,650,000.00
2	ADVANCED DISSECTING MICROSCOPE	400,000.00
3	AUTOCLAVE (STERILIZER)	200,000.00
4	DIGITAL WATER FLOW METER	600,000.00
5	GRAB TYPE BOTTOM SAMPLING DREDGE	150,000.00
	TOTAL ABC	3,000,000.00

- The *Carlos Hilado Memorial State University* now invites bids for the above Procurement Project. Delivery of the Services is required by **Sixty (60) calendar days after receipt of Notice to Proceed**. Bidders should have completed, within two (2) years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II. Instructions to Bidders.
- Bidding will be conducted through open competitive bidding procedures using the non-discretionary *pass/fail* criterion as specified in the 2016 Revised Implementing Rules and Regulations (IRR) of Republic Act 9184.

Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA 5183.

- Prospective bidders may obtain further information from CARLOS HILADO MEMORIAL STATE UNIVERSITY and inspect the Bidding Documents at the address given below during 8:00 A.M. to 5:00 P.M.

The BAC Secretariat
Brgy. Zone I, Mabini St.,
Talisay City, Negros Occidental

- A complete set of Bidding Documents may be acquired by interested Bidders on **January 25 – February 15, 2024 (8:30A.M.)** from the given address and website(s) below and upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, in the maximum amount of **Five Thousand Pesos (Php 5,000.00) Only**. The Procuring Entity shall allow the bidder to present its proof of payment for the fees.



✉ bac.sec@chmsc.edu.ph
☎ (034) 712 0005 local 142
🌐 chmsc.edu.ph

GREEN CHMSU EXCELLENCE!

Excellence • Compassion • Environmentalism • Love of Country • Social Responsibility • Integrity • Discipline • Resilience



Carlos Hilado Memorial State University

Alijis Campus • Binalbagan Campus • Fortune Towne Campus • Talisay (Main) Campus

A leading GREEN institution of higher learning in the global community by 2030

Bids and Awards Committee

ITEM NO.	DESCRIPTION	ABC PER ITEM	COST OF BIDDING DOCUMENTS
1	ABSORBANCE/FLOURESCENE SPECTROSCOPY	1,650,000.00	5,000.00
2	ADVANCED DISSECTING MICROSCOPE	400,000.00	500.00
3	AUTOCCLAVE (STERILIZER)	200,000.00	500.00
4	DIGITAL WATER FLOW METER	600,000.00	1,000.00
5	GRAB TYPE BOTTOM SAMPLING DREDGE	150,000.00	500.00

- The CHMSU will hold a **Pre-Bid conference on 9:30 A.M., February 1, 2024** at **Bidding Room, 2/F Supply and Property Management Bldg., Brgy. Zone 1, Mabini St., Talisay City, Negros Occidental** and or through video conferencing or webcasting via **Zoom Meeting with ID No. 797 389 7583, Meeting Password 020124**, which shall be open to prospective bidders.
- Bids must be duly received by the BAC Secretariat through (i) manual submission at the **Bidding Room, 2/F Supply and Property Management Bldg., Brgy. Zone 1, Mabini St., Talisay City, Negros Occidental**, (ii) online or electronic submission at bac.sec@chmsu.edu.ph or (iii) both on or before **9:00 A.M., February 15, 2024**. Late bids shall not be accepted.

Bids may be submitted through electronic mail to bac.sec@chmsu.edu.ph provided that the bidding documents are compressed into two (2) separate archived folders (zip or rar format) and which each folder shall be labelled as "First Envelope_Name of Company_Project Reference Number" and "Second Envelope_Name of Company_Project Reference Number" and each is uniquely password-protected;

Bidders must submit the printed copies (Original, Copy 1 and Copy 2) of their bidding documents within 3 calendar days from bid opening.

- All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in ITB Clause 14.
- Bid opening shall be on **9:30 A.M., February 15, 2024**, at the **Bidding Room, 2/F Supply and Property Management Bldg., Brgy. Zone 1, Mabini St., Talisay City, Negros Occidental** through **Zoom Meeting ID No. 797 389 7583, Meeting Password 021524**. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.
- The Carlos Hilado Memorial State University reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA 9184, without thereby incurring any liability to the affected bidder or bidders.
- For further information, please refer to:

MS. LIGAYA E. FUENTES
Head-BAC Secretariat
Carlos Hilado Memorial State University
Brgy. Zone 1, Mabini St.,
Talisay City, Negros Occidental
Email Add.: bac.sec@chmsu.edu.ph
Telephone Nos.: (034) 454-0529; 454-0584 local 142
Mobile No.: 0939-9296624
Website: chmsu.edu.ph



 bac.sec@chmsc.edu.ph
 (034) 712 0005 local 142
 chmsc.edu.ph

GREEN CHMSU ExCELSIOR!

Excellence • Compassion • Environmentalism • Love of Country • Social Responsibility • Integrity • Openness • Resilience



Carlos Hilado Memorial State University

Alijis Campus • Binalbagan Campus • Fortune Towne Campus • Talisay (Main) Campus

A leading GREEN institution of higher learning in the global community by 2030

Bids and Awards Committee

12. You may visit the following websites:

For downloading of Bidding Documents: chmsu.edu.ph

For online bid submission: bac.sec@chmsu.edu.ph


ANDREW EUSEBIO S. TAN, Ph.D.
BAC Chairperson



 bac.sec@chmsc.edu.ph
 (034) 712 0005 local 142
 chmsc.edu.ph

GREEN CHMSU ExCELSIOR!

Excellence • Compassion • Environmentalism • Love of Country • Social Responsibility • Integrity • Openness • Resilience

Section II. Instructions to Bidders

Notes on the Instructions to Bidders

This Section on the Instruction to Bidders (ITB) provides the information necessary for bidders to prepare responsive bids, in accordance with the requirements of the Procuring Entity. It also provides information on bid submission, eligibility check, opening and evaluation of bids, post-qualification, and on the award of contract.

1. Scope of Bid

The Procuring Entity, CARLOS HILADO MEMORIAL STATE UNIVERSITY wishes to receive Bids for the **PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUCTIONAL AND RESEARCH OF THE COLLEGE OF FISHERIES – BINALBAGAN CAMPUS** with identification number **CHMSU 24-011-0124-G**.

The Procurement Project (referred to herein as “Project”) is composed of **Five (5) Items**, the details of which are described in Section VII (Technical Specifications).

2. Funding Information

2.1. The GOP through the source of funding as indicated below for 2024 in the amount of **THREE MILLION PESOS & 00/100 (Php3,000,000.00) ONLY**.

2.2. The source of funding is:

NGA, the General Appropriations Act or Special Appropriations.

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex “I” of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. Foreign ownership exceeding those allowed under the rules may participate pursuant to:
 - i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
 - ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
 - iii. When the Goods sought to be procured are not available from local suppliers; or
 - iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:
 - a. For the procurement of Non-expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

7. Subcontracts

The Procuring Entity has prescribed that:

Subcontracting is not allowed.

8. Pre-Bid Conference

The Procuring Entity will hold a **Pre-Bid Conference** for this Project on **February 1, 2024, 9:30 A.M.** at the Bidding Rm., 2/F Supply and Property Management Bldg., CHMSU Talisay Campus.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within two (2) years prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

- 11.5. Financial proposals for single or multi-year Framework Agreement shall be submitted before the deadline of submission of bids as prescribed in the **IB**. For multi-year Framework Agreement, evaluation of the financial proposal during this stage is for purposes of determining eligibility and whether or not such financial proposal is within the ABC.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:

- a. For Goods offered from within the Procuring Entity's country:

- i. The price of the Goods quoted EXW (ex-works, ex-factory, ex-warehouse, ex-showroom, or off-the-shelf, as applicable);
- ii. The cost of all customs duties and sales and other taxes already paid or payable;
- iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
- iv. The price of other (incidental) services, if any, listed in e.

- b. For Goods offered from abroad:

- i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
- ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications)**.

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.

13.2. Payment of the contract price shall be made in:

- a. Philippine Pesos.

14. Bid Security

14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.

14.2. The Bid and bid security shall be valid until **June 14, 2024**. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

16.1. The Bidders shall submit on the specified date and time and either at its physical address or through online submission as indicated in paragraph 7 of the **IB**.

17. Opening and Preliminary Examination of Bids

17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

- 17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

- 18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "*passed*," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as follows:
- One Project having several items grouped into several lots, which shall be awarded as separate contracts per lot.
- 19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid. The Bidder shall

submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**.

21. Signing of the Contract

- 21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

Section III. Bid Data Sheet

Notes on the Bid Data Sheet

The Bid Data Sheet (BDS) consists of provisions that supplement, amend, or specify in detail, information, or requirements included in the ITB found in Section II, which are specific to each procurement.

This Section is intended to assist the Procuring Entity in providing the specific information in relation to corresponding clauses in the ITB and has to be prepared for each specific procurement.

The Procuring Entity should specify in the BDS information and requirements specific to the circumstances of the Procuring Entity, the processing of the procurement, and the bid evaluation criteria that will apply to the Bids. In preparing the BDS, the following aspects should be checked:

- a. Information that specifies and complements provisions of the ITB must be incorporated.
- b. Amendments and/or supplements, if any, to provisions of the ITB as necessitated by the circumstances of the specific procurement, must also be incorporated.

Bid Data Sheet

ITB Clause																
5.3	<p>For this purpose, contracts similar to the Project shall be:</p> <p style="margin-left: 40px;">a. PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUCTIONAL AND RESEARCH OF THE COLLEGE OF FISHERIES – BINALBAGAN CAMPUS</p> <p style="margin-left: 40px;">b. completed within two (2) years prior to the deadline for the submission and receipt of bids.</p>															
7.1	Subcontracting is not allowed.															
12	The price of the Goods shall be quoted DDP or the applicable International Commercial Terms (INCOTERMS) for this Project.															
14.1	<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <p style="margin-left: 40px;">a. The amount of not less than 60,000.00 (2%) of ABC, if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</p> <p style="margin-left: 40px;">b. The amount of not less than 150,000.00 (5%) of ABC, if bid security is in Surety Bond.</p>															
19.3	<table style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td style="width: 10%;">Item 1</td> <td style="width: 70%;">Absorbance/Flourescene Spectroscopy</td> <td style="width: 20%; text-align: right;">1,650,000.00</td> </tr> <tr> <td>Item 2</td> <td>Advanced Dissecting Microscope</td> <td style="text-align: right;">400,000.00</td> </tr> <tr> <td>Item 3</td> <td>Autoclave (Sterilizer)</td> <td style="text-align: right;">200,000.00</td> </tr> <tr> <td>Item 4</td> <td>Digital Water Flow Meter</td> <td style="text-align: right;">600,000.00</td> </tr> <tr> <td>Item 5</td> <td>Grab Type Bottom Sampling Dredge and Observation Head</td> <td style="text-align: right; vertical-align: bottom;">150,000.00</td> </tr> </tbody> </table>	Item 1	Absorbance/Flourescene Spectroscopy	1,650,000.00	Item 2	Advanced Dissecting Microscope	400,000.00	Item 3	Autoclave (Sterilizer)	200,000.00	Item 4	Digital Water Flow Meter	600,000.00	Item 5	Grab Type Bottom Sampling Dredge and Observation Head	150,000.00
Item 1	Absorbance/Flourescene Spectroscopy	1,650,000.00														
Item 2	Advanced Dissecting Microscope	400,000.00														
Item 3	Autoclave (Sterilizer)	200,000.00														
Item 4	Digital Water Flow Meter	600,000.00														
Item 5	Grab Type Bottom Sampling Dredge and Observation Head	150,000.00														
20.2	Latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS).															
21.2	No further instructions.															

Section IV. General Conditions of Contract

Notes on the General Conditions of Contract

The General Conditions of Contract (GCC) in this Section, read in conjunction with the Special Conditions of Contract in Section V and other documents listed therein, should be a complete document expressing all the rights and obligations of the parties.

Matters governing performance of the Supplier, payments under the contract, or matters affecting the risks, rights, and obligations of the parties under the contract are included in the GCC and Special Conditions of Contract.

Any complementary information, which may be needed, shall be introduced only through the Special Conditions of Contract.

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

2. Advance Payment and Terms of Payment

2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.

2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the SCC.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the SCC, **Section IV (Technical Specifications)** shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

- 6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.
- 6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

Section V. Special Conditions of Contract

Notes on the Special Conditions of Contract

Similar to the BDS, the clauses in this Section are intended to assist the Procuring Entity in providing contract-specific information in relation to corresponding clauses in the GCC found in Section IV.

The Special Conditions of Contract (SCC) complement the GCC, specifying contractual requirements linked to the special circumstances of the Procuring Entity, the Procuring Entity's country, the sector, and the Goods purchased. In preparing this Section, the following aspects should be checked:

- a. Information that complements provisions of the GCC must be incorporated.
- b. Amendments and/or supplements to provisions of the GCC as necessitated by the circumstances of the specific purchase, must also be incorporated.

However, no special condition which defeats or negates the general intent and purpose of the provisions of the GCC should be incorporated herein.

Special Conditions of Contract

GCC Clause	
1	<p>Delivery and Documents –</p> <p>For purposes of the Contract, “EXW,” “FOB,” “FCA,” “CIF,” “CIP,” “DDP” and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:</p> <p><i>[For Goods supplied from abroad, state:]</i> “The delivery terms applicable to the Contract are DDP delivered Carlos Hilado Memorial State College, Talisay City, Negros Occidental . In accordance with INCOTERMS.”</p> <p><i>[For Goods supplied from within the Philippines, state:]</i> “The delivery terms applicable to this Contract are delivered at Carlos Hilado Memorial State College, Talisay City, Negros Occidental Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination.”</p> <p>Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).</p> <p>For purposes of this Clause the Procuring Entity’s Representative at the Project Site is Mr. Philip C. Estacion, Supply Officer.</p> <p>Incidental Services –</p> <p>The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements:</p> <ol style="list-style-type: none"> a. performance or supervision of on-site assembly and/or start-up of the supplied Goods; b. furnishing of tools required for assembly and/or maintenance of the supplied Goods; c. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods; d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and

- e. training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.

The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

Spare Parts –

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
 - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
 - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in **Section VI (Schedule of Requirements)** and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of one (1) year.

Spare parts or components shall be supplied as promptly as possible, but in any case, within fifteen (15) days of placing the order.

	<p>Packaging –</p> <p>The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods’ final destination and the absence of heavy handling facilities at all points in transit.</p> <p>The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.</p> <p>The outer packaging must be clearly marked on at least four (4) sides as follows:</p> <p>Name of the Procuring Entity Name of the Supplier Contract Description Final Destination Gross weight Any special lifting instructions Any special handling instructions Any relevant HAZCHEM classifications</p>
	<p>A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.</p> <p>Transportation –</p> <p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.</p> <p>Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.</p>

	<p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.</p> <p>The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.</p> <p>Intellectual Property Rights –</p> <p>The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.</p>
2.2	<p>The terms of payment shall be as follows:</p> <p>The terms of payment : 30 days</p>
4	<p>The inspections and tests that will be conducted are:</p> <p>Inspection as to the Quality, Quantity, and Specifications of the items delivered.</p>

Section VI. Schedules of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item #	Description	Quantity	Total	Delivered, Weeks/Months
1	ABSORBANCE/FLOURESCENCE SPECTROSCOPY	1	1	
	Comprising of: 1 set Flame-Chem-Vis-Nir, 1 pc Oceanview Software/ 1 License, 2 Installations, 1 pack (100 pcs) CVD-VIS1M Plastic Cuvettes			
	Technical Specifications: Spectral range: 350-1000 nm, Spectrometer (preconfigured)			
	FLAME-S-VIS-NIR-ES, Detector: Sony ILX511B linear silicon CCD array. Entrance slit: 25 um slit, Optical resolution: -1.5 nm(FWHM), Order-sorting filter: 350-1000 nm, Focus lens for enhanced sensitivity: Yes, Signal-to-noise ratio: 250:1			
	Dynamic Range: 1300:1 for a single acquisition, Integration Time: 1 ms-65 seconds, Stray light <0.05% at 600 nm, (INTEGRATED SAMPLING SYSTEMS) FLAME-DA-CUV-VIS-NIR, Light source: Tungsten halogen with a violet LED (-390-1000 nm) Light source features: I ² c adjustment of the halogen and LED bulbs, Bulb lifetime: 2000 hours (tungsten halogen); 45000 hours (LED), Cuvette holder: Accepts 1 cm, Pathlength cuvettes, Cuvette holder features: Improved collection optics			
	After Sales Support: Product Demonstration, Installation and Calibration			
2	ADVANCED DISSECTING MICROSCOPE	2	2	
	Total Magnification: 10X-20X-40X (10X Eyepiece) Working distance: 90mm.Focus: Focusing knobs on stand. BSERVATION HEAD : Binocular head Inclined at 45 rotatable 3 60 lockable at any angle. Both eyepiece tubes focusable to nmatch objectives tube length, interpupillary distance adjustments provided. EYEPIECE : Lockable, anti-fungal KEOWA Extra Wide Field HKW 10x / FOV 18mm(PAIRED) A.R. Coated eyepiece with high eyepoint and foldable eyeguards offers maximum resolution and high colour nfidelity and perfectly designed for a FLAT FIELD. H 6x (PAIRED) eyepiece EXTRA NOSEPIECE : Quadruple nosepiece precisely built- on ball bearing, provides lifelong smooth rotation through a built- nin positive click stop and rubber grip. Totally, dust proof design. OBJECTIVES : KEOWA DIN size RMS S-Plan Achromatic objectives 4x/NA 0.10, 10x/NA 0.25, 40x/NA 0.65 (SL) n and 100x/NA 1.25 (SL) oil immersion, anti-fungal. MECHANICAL STAGE : Double Plated Rectangular stage (Ball Bearing Type) with a surface of 120 mm x 140mm having low npositioned co-axial controls. Both X&Y axis having graduated scale & verniersreading to 0.1mm. CO-AXIAL FOCUSING : Co-axial fine and coarse focusing control with ball bearing guideways based on four stage gear reduction system, provides the degree of reduction sensitivity, free from any BACK LASH & large nknobs provide a soft and sophisticated touch to the microscope. TENSION CONTROL RING : Enable you to adjust the tension on coarse focussing knobs. CONDENSER : A focusable Abbe two lens type NA 1.25 condenser with iris diaphragm and swing out filter holder with nBlue filter. Rack & pinion movement on metal guides. EXTRA LARGE BASE : Single mold sturdy stand with anti rust materials ILLUMINATION : Built-in base illumination special quality HALOGEN bulb 6V-20W and intensity controlled power supply. OPTIONAL ACCESSORIES : High resolution digital camera, Image Analysis Software.			
	After Sales Support: Product Demonstration, Installation and Calibration			
3	AUTOCLAVE (STERILIZER)	1	1	
	Temp Range Max 134°C (273°F) Using Press 0.12MPa Time Controller			

	Microprocessor Digital PID Controller Sterling Time 00.00 to 99HR 59MIN (MIN SEC) selectable			
	Pressure Gauge Mechanical type 0 to 0.4 Mpa Sensor PT-100 Capa 60L			
	Dimension Internal $\Phi 344 \times 650$ (H)mm / $\Phi 13.54 \times 25.59$ Dimension			
	External (WxDxH) 550x680x1090mm / 21.65"x26.77"x42.91" Net			
	Weight 110Kg Material Internal Stainless Steel Material External Steel			
	Plate with Powder Heating Coated Heater 3KW Electrical Requirements			
	120V, 60Hz, 1 Φ 25.0A Cat.No. AC60M120 Electrical Requirements			
	230V, 60Hz, 1 Φ , After Sales Support: Product Demonstration			
4	DIGITAL WATER FLOW METER	2	2	
	Has a flow sensor with a magnetic multi-bladed spinning rotor (impeller) mounted inside at right angles to the flow. The rotational speed of the impeller is proportional to the velocity of fluid. The pulse generated is transmitted and converted into a volumetric measurement. Our digital flow meters are individually calibrated to give reliable flow measurement. Digital Flow Meter measuring flow rate; 0.05 m ³ /hr to 500 m ³ /hr Physical dimensions; 105 x 105 x 120mm, Cut-out size; 92 x 92mm Enclosure; Abs weather proof IP-65, Mounting; Field / Panel Supply power; 230 VA.C./ 110 VA.C./ 24 V (200mA) D.C. Max. Total Count: 0-999999.9 m ³ / 0-9999999 liters, Accuracy; Calibration/set point; Using front panel keypad After Sales Support: Product Demonstration, Installation and Calibration			
5	GRAB TYPE BOTTOM SAMPLING DREDGE	1	1	
	Ideal for sampling sand, silt and non-cohesive material Hinged scissor jaws that close when the sampler is lifted Designed for fast deployment & reduced sample loss Spring-loaded closing mechanisms Collects samples as they were in-situ to make sediment profile descriptions possible Made from strong, stainless steel. 2.3 kg (5 lb.) Sand Silt Dredge; 6.8 kg (15 lb.) & 11.3 kg (25 lb.) Bottom Dredges; E429 - The 2.3kg, 6.8 kg and 11.3 kg bottom dredges After Sales Support: Product Demonstration			
	-x-x-x-x-x-x-x-x-x-x-nothing follows-x-x-x-x-x-x-x-x-x-x-			
	TOTAL ABC = Php 3,000,000.00			
	PR # 24-006-1123 11-23-2023/I. Perante			
	MDS 07-101-24-01(1) 01-02-2024			
	Note: All items should be delivered within Sixty (60) calendar days after receipt of notice to proceed.			

Section VII. Technical Specifications

Item	Specification	Statement of Compliance
		<p>Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a.2) and/or GCC Clause 2.1(a.2).</p>
1	<p>ABSORBANCE/FLOURESCENCE SPECTROSCOPY Comprising of: 1 set Flame-Chem-Vis-Nir, 1 pc Oceanview Software/ 1 License, 2 Installations, 1 pack (100 pcs) CVD-VIS1M Plastic Cuvettes Technical Specifications: Spectral range: 350-1000 nm, Spectrometer (preconfigured) FLAME-S-VIS-NIR-ES, Detector: Sony ILX511B linear silicon CCD array. Entrance slit: 25 um slit, Optical resolution: -1.5 nm(FWHM), Order-sorting filter:350-1000 nm, Focus lens for enhanced sensitivity: Yes, Signal-to-noise ratio: 250:1 Dynamic Range: 1300:1 for a single acquisition, Integration Time: 1 ms-65 seconds, Stray light <0.05% at 600 nm, (INTEGRATED SAMPLING SYSTEMS) FLAME-DA-CUV-VIS-NIR, Light source: Tungsten halogen with a violet LED (-390-1000 nm) Light source features: I²c adjustment of the halogen and LED bulbs, Bulb lifetime: 2000 hours (tungsten halogen); 45000 hours (LED), Cuvette holder: Accepts 1 cm, Pathlength cuvettes, Cuvette holder features: Improved collection optics After Sales Support: Product Demonstration, Installation and Calibration</p>	
2	<p>ADVANCED DISSECTING MICROSCOPE Total Magnification: 10X-20X-40X (10X Eyepiece) Working distance: 90mm.Focus: Focusing knobs on stand. BSRVATION HEAD : Binocular head Inclined at 45 rotatable 3 60 lockable at any angle. Both eyepiece tubes focusable to nmatch objectives tube length, interpupillary distance adjustments provided. EYEPIECE : Lockable, anti-fungal KEOWA Extra Wide Field HKW 10x /FOV 18mm(PAIRED) A.R. Coated eyepiece with high eyepoint and foldable eyeguards offers maximum resolution and high colour nfidely and perfectly</p>	

	<p>designed for a FLAT FIELD. H 6x (PAIRED) eyepiece</p> <p>EXTRA NOSEPIECE : Quadruple nosepiece precisely built- on ball bearing, provides lifelong smooth rotation through a built- in positive click stop and rubber grip. Totally, dust proof design.</p> <p>OBJECTIVES : KEOWA DIN size RMS S-Plan Achromatic objectives 4x/NA 0.10, 10x/NA 0.25, 40x/NA 0.65 (SL) n and 100x/NA 1.25 (SL) oil immersion, anti-fungal.</p> <p>MECHANICAL STAGE : Double Plated Rectangular stage (Ball Bearing Type) with a surface of 120 mm x 140mm having low npositioned co-axial controls. Both X&Y axis having graduated scale & verniersreading to 0.1mm.</p> <p>CO-AXIAL FOCUSING : Co-axial fine and coarse focusing control with ball bearing guideways based on four stage gear reduction system, provides the degree of reduction sensitivity, free from any BACK LASH & large nknobs provide a soft and sophisticated touch to the microscope.</p> <p>TENSION CONTROL RING : Enable you to adjust the tension on coarse focussing knobs.</p> <p>CONDENSER : A focusable Abbe two lens type NA 1.25 condenser with iris diaphragm and swing out filter holder with nBlue filter. Rack & pinion movement on metal guides.</p> <p>EXTRA LARGE BASE : Single mold sturdy stand with anti rust materials, ILLUMINATION : Built-in base illumination special quality HALOGEN bulb 6V-20W and intensity controlled power supply.</p> <p>OPTIONAL ACCESSORIES : High resolution digital camera, Image Analysis Software.</p> <p>After Sales Support: Product Demonstration, Installation and Calibration</p>	
3	<p>AUTOCLAVE (STERILIZER)</p> <p>Temp Range Max 134°C (273°F) Using Press 0.12MPa Time Controller Microprocessor Digital PID Controller Sterling</p> <p>Time 00.00 to 99HR 59MIN (MIN SEC) selectable</p> <p>Pressure Gauge Mechanical type 0 to 0.4 Mpa Sensor PT-100</p> <p>Capa 60L, Dimension Internal Φ344x650(H)mm / Φ13.54"x25.59</p> <p>Dimension External (WxDxH) 550x680x1090mm / 21.65"x26.77"x42.91" Net Weight 110Kg Material Internal Stainless Steel Material External Steel Plate with Powder Heating</p> <p>Coated Heater 3KW Electrical Requirements 120V, 60Hz, 1Φ 25.0A Cat.No. AC60M120 Electrical Requirements 230V, 60Hz, 1Φ, After Sales Support: Product Demonstration</p>	
4	<p>DIGITAL WATER FLOW METER</p> <p>Has a flow sensor with a magnetic multi-bladed spinning rotor (impeller) mounted inside at right angles to the flow. The rotational speed of the impeller is proportional to the velocity of fluid. The pulse generated is transmitted and converted into a volumetric measurement. Our digital flow meters are individually calibrated to give reliable flow measurement. Digital Flow Meter measuring flow rate; 0.05 m³/hr to 500 m³/hr</p> <p>Physical dimensions; 105 x 105 x 120mm, Cut-out size; 92 x 92mm Enclosure; Abs weather proof IP-65, Mounting; Field / Panel, Supply power; 230 VA.C./ 110 VA.C./ 24 V (200mA) D.C. Max. Total Count: 0-999999.9 m³ / 0-9999999 liters,</p> <p>Accuracy; Calibration/set point; Using front panel keypad</p> <p>After Sales Support: Product Demonstration, Installation and Calibration</p>	
5	<p>GRAB TYPE BOTTOM SAMPLING DREDGE</p> <p>Ideal for sampling sand, silt and non-cohesive material</p> <p>Hinged scissor jaws that close when the sampler is lifted</p> <p>Designed for fast deployment & reduced sample loss</p> <p>Spring-loaded closing mechanisms</p> <p>Collects samples as they were in-situ to make sediment profile descriptions possible</p> <p>Made from strong, stainless steel.</p>	

	2.3 kg (5 lb.) Sand Silt Dredge; 6.8 kg (15 lb.) & 11.3 kg (25 lb.)	
	Bottom Dredges; E429 - The 2.3kg, 6.8 kg and 11.3 kg bottom dredges	
	After Sales Support: Product Demonstration	
	-x-x-x-x-x-x-x-x-x-x-nothing follows-x-x-x-x-x-x-x-x-x-x-	
	TOTAL ABC = Php 3,000,000.00	
	PR # 24-006-1123 11-23-2023/I. Perante	
	MDS 07-101-24-01(1) 01-02-2024	



**CARLOS HILADO MEMORIAL STATE UNIVERSITY
BIDS AND AWARDS COMMITTEE**

Talisay City, Negros Occidental
Tel. Nos.: (034) 454-0529; 454-0584 Local 142
Mobile No.: 0939-9296624

Project Reference No: **CHMSU 24-011-0124-G**
 Name of the Project: **PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUMENTAL AND RESEARCH OF THE COLLEGE OF FISHERIES**
 Location of the Project: **BINALBAGAN CAMPUS**

BILL OF QUANTITIES

Date: _____

Item No.	Qty	Unit	ARTICLES and DESCRIPTION	ABC / UNIT	Unit Price	Total Price
1	1	unit	ABSORBANCE/FLOURESCENCE SPECTROSCOPY	1,650,000.00		
			Comprising of: 1 set Flame-Chem-Vis-Nir, 1 pc Oceanview Software/ 1 License, 2 Installations, 1 pack (100 pcs) CVD-VIS1M Plastic Cuvettes			
			Technical Specifications: Spectral range: 350-1000 nm, Spectrometer (preconfigured)			
			FLAME-S-VIS-NIR-ES, Detector: Sony ILX511B linear silicon CCD array.			
			Entrance slit: 25 um slit, Optical resolution: -1.5 nm(FWHM), Order-sorting filter: 350-1000 nm, Focus lens for enhanced sensitivity: Yes, Signal-to-noise ratio: 250:1			
			Dynamic Range: 1300:1 for a single acquisition, Integration Time: 1 ms-65 seconds, Stray light <0.05% at 600 nm, (INTEGRATED SAMPLING SYSTEMS) FLAME-DA-CUV-VIS-NIR, Light source:			
			Tungsten halogen with a violet LED (-390-1000 nm) Light source features: I ² c adjustment of the halogen and LED bulbs, Bulb lifetime: 2000 hours (tungsten halogen); 45000 hours (LED), Cuvette holder: Accepts 1 cm, Pathlength cuvelles, Cuvette holder features:			
			Improved collection optics			
			After Sales Support: Product Demonstration, Installation and Calibration			
2	2	unit	ADVANCED DISSECTING MICROSCOPE	200,000.00		
			Total Magnification: 10X-20X-40X (10X Eyepiece)			
			Working distance: 90mm.Focus: Focusing knobs on stand.			
			BSERVATION HEAD : Binocular head Inclined at 45 rotatable 3 60 lockable at any angle. Both eyepiece tubes focusable to nmatch objectives tube length, interpupillary distance adjustments provided.			
			EYEPIECE : Lockable, anti-fungal KEOWA Extra Wide Field HKW 10x / FOV 18mm(PAIRED) A.R. Coated eyepiece with high eyepoint and foldable eyeguards offers maximum resolution and high colour nfidely and perfectly designed for a FLAT FIELD. H 6x (PAIRED) eyepiece			
			EXTRA NOSEPIECE : Quadruple nosepiece precisely built- on ball bearing, provides lifelong smooth rotation through a built- nin positive click stop and rubber grip. Totally, dust proof design.			
			OBJECTIVES : KEOWA DIN size RMS S-Plan Achromatic objectives 4x/NA 0.10, 10x/NA 0.25, 40x/NA 0.65 (SL) n and 100x/NA 1.25 (SL) oil immersion, anti-fungal.			
			MECHANICAL STAGE : Double Plated Rectangular stage (Ball Bearing Type) with a surface of 120 mm x 140mm having low npositioned co-axial controls. Both X&Y axis having graduated scale & verniersreading to 0.1mm.			

Signature of Bidder _____
 Name of Firm _____
 Date _____



CARLOS HILADO MEMORIAL STATE UNIVERSITY
BIDS AND AWARDS COMMITTEE
 Talisay City, Negros Occidental
 Tel. Nos.: (034) 454-0529; 454-0584 Local 142
 Mobile No.: 0939-9296624

Project Reference No: **CHMSU 24-011-0124-G**
 Name of the Project: **PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUCTIONAL AND RESEARCH OF THE COLLEGE OF FISHERIES**
 Location of the Project: **BINALBAGAN CAMPUS**

page 2 of 3

BILL OF QUANTITIES

Date: _____

Item No.	Qty	Unit	ARTICLES and DESCRIPTION	ABC / UNIT	Unit Price	Total Price
			CO-AXIAL FOCUSING : Co-axial fine and coarse focusing control with ball bearing guideways based on four stage gear reduction system, provides the degree of reduction sensitivity, free from any BACK LASH & large knobs provide a soft and sophisticated touch to the microscope.			
			TENSION CONTROL RING : Enable you to adjust the tension on coarse focussing knobs.			
			CONDENSER : A focusable Abbe two lens type NA 1.25 condenser with iris diaphragm and swing out filter holder with nBlue filter. Rack & pinion movement on metal guides.			
			EXTRA LARGE BASE : Single mold sturdy stand with anti rust materials			
			ILLUMINATION : Built-in base illumination special quality			
			HALOGEN bulb 6V-20W and intensity controlled power supply.			
			OPTIONAL ACCESSORIES : High resolution digital camera, Image Analysis Software.			
			After Sales Support: Product Demonstration, Installation and Calibration			
3	1	unit	AUTOCLAVE (STERILIZER)	200,000.00		
			Temp Range Max 134°C (273°F) Using Press 0.12MPa Time Controller			
			Microprocessor Digital PID Controller Sterling Time 00.00 to 99HR 59MIN (MIN SEC) selectable			
			Pressure Gauge Mechanical type 0 to 0.4 Mpa Sensor PT-100 Capa 60L			
			Dimension Internal Φ344x650(H)mm / Φ13.54"x25.59 Dimension			
			External (WxDxH) 550x680x1090mm / 21.65"x26.77"x42.91" Net			
			Weight 110Kg Material Internal Stainless Steel Material External Steel			
			Plate with Powder Heating Coated Heater 3KW Electrical Requirements			
			120V, 60Hz, 1Φ 25.0A Cat.No. AC60M120 Electrical Requirements			
			230V, 60Hz, 1Φ, After Sales Support: Product Demonstration			
4	2	unit	DIGITAL WATER FLOW METER	300,000.00		
			Has a flow sensor with a magnetic multi-bladed spinning rotor (impeller) mounted inside at right angles to the flow. The rotational speed of the impeller is proportional to the velocity of fluid. The pulse generated is transmitted and converted into a volumetric measurement.			
			Our digital flow meters are individually calibrated to give reliable flow measurement.			
			Digital Flow Meter measuring flow rate; 0.05 m ³ /hr to 500 m ³ /hr			
			Physical dimensions; 105 x 105 x 120mm, Cut-out size; 92 x 92mm			
			Enclosure; Abs weather proof IP-65, Mounting; Field / Panel			
			Supply power; 230 VA.C./ 110 VA.C./ 24 V (200mA) D.C.			

Signature of Bidder _____
 Name of Firm _____
 Date _____



CARLOS HILADO MEMORIAL STATE UNIVERSITY
BIDS AND AWARDS COMMITTEE
 Talisay City, Negros Occidental
 Tel. Nos.: (034) 454-0529; 454-0584 Local 142
 Mobile No.: 0939-9296624

Project Reference No: **CHMSU 24-011-0124-G**
 Name of the Project: **PROCUREMENT OF VARIOUS EQUIPMENT OF CAPTURE FISHERIES LABORATORY FOR INSTRUCTIONAL AND RESEARCH OF THE COLLEGE OF FISHERIES**
 Location of the Project: **BINALBAGAN CAMPUS**

page 3 of 3

BILL OF QUANTITIES

Date: _____

Item No.	Qty	Unit	ARTICLES and DESCRIPTION	ABC / UNIT	Unit Price	Total Price
			Max. Total Count: 0-999999.9 m ³ / 0-9999999 liters, Accuracy;			
			Calibration/set point; Using front panel keypad			
			After Sales Support: Product Demonstration, Installation and			
			Calibration			
5	1	unit	GRAB TYPE BOTTOM SAMPLING DREDGE	150,000.00		
			Ideal for sampling sand, silt and non-cohesive material			
			Hinged scissor jaws that close when the sampler is lifted			
			Designed for fast deployment & reduced sample loss			
			Spring-loaded closing mechanisms			
			Collects samples as they were in-situ to make sediment profile			
			descriptions possible			
			Made from strong, stainless steel.			
			2.3 kg (5 lb.) Sand Silt Dredge; 6.8 kg (15 lb.) & 11.3 kg (25 lb.)			
			Bottom Dredges; E429 - The 2.3kg, 6.8 kg and 11.3 kg bottom dredges			
			After Sales Support: Product Demonstration			
			-X-			
			TOTAL ABC = Php 3,000,000.00			
			PR # 24-006-1123 11-23-2023/I. Perante			
			MDS 07-101-24-01(1) 01-02-2024			

Signature of Bidder _____
 Name of Firm _____
 Date _____